

PATIENT ALERT

Abbott FreeStyle Libre 3 and FreeStyle Libre 3 Plus Sensors Recall December 4, 2025

The manufacturer of FreeStyle Libre 3 and FreeStyle Libre 3 Plus Sensors has issued a recall for some of its sensors. They have recently identified that certain lots may provide incorrect low glucose readings. A few of affected lots were dispensed through Gila River Health Care Hu Hu Kam Memorial Hospital and Komatke Health Center Pharmacies. To determine if sensors are potentially affected, patients can visit www.FreeStyleCheck.com and select "CONFIRM SENSOR SERIAL NUMBER." Patients can also call Life Center 520-562-7940 or Pharmacy 602-528-1303 during business hours (see below). The serial number can be found in the app or reader and also on the label on the bottom of the sensor applicator or carton. Those using the affected sensors are strongly advised to discontinue use and dispose of the sensors. Any unused affected sensor(s) can be replaced at Gila River Health Care pharmacies.

This issue does not apply to any other Libre sensors, Accu-Chek strips or Dexcom sensors.

If you have any questions please contact either:

Pharmacy

602-528-1303

Monday - Saturday 8:00am - 8:00pm

Sunday 12:00pm - 8:00pm

Life Center

520-562-7940

Monday - Friday 8:00am - 5:00pm